


An Easy Guide to TENS Pain Relief

A book of indications, electrode placement & suggested protocol settings

ElectroMedical Solutions 1715 E. Bay Dr. Largo, FL 33771 Toll Free 866-243-TENS (8367) Toll Free Fax 866-242-TENS (8367) www.electromedicalsolutions.com



LOW BACK PAIN

➢ Setting

MODE: C Mode

PULSE WIDTH: 260

PULSE RATE: 50-80Hz


OUTPUT: Adjust to the most comfortable intensity that the muscles feel.


Image: Alternate Placement

Treatment Session

24 hours is available until initial relief.







HIP NEURALGIA

➢ Setting

MODE: M Mode

PULSE WIDTH: 150-260

PULSE RATE: 80-120Hz


OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available for the first 4 days.

4 hours per day thereafter.





PHANTOM LIMB, LOWER EXTREMITY


➢ Setting

MODE: C Mode or M Mode

PULSE WIDTH: 160 - 200

PULSE RATE: 50 - 100Hz

OUTPUT: Adjust to the most comfortable intensity level.




Treatment Session

24 hours everyday for a week is available.



SCIATICA



➢ Setting

MODE: M Mode


PULSE WIDTH: 260


PULSE RATE: 150Hz


OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session









BICIPITAL TENDONITIS

➢ Setting

MODE: M Mode


PULSE WIDTH: 150 - 160


PULSE RATE: 50Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session







TEMPORAL MANDIBULAR JOINT PAIN (TMJ)

➢ Setting

MODE: M Mode

PULSE WIDTH: 220

PULSE RATE: 10Hz


OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available for the first 2 days.

6 hours daily thereafter.





SHOULDER PAIN


➢ Setting

MODE: M Mode

PULSE WIDTH: 260

PULSE RATE: 80 - 100Hz


OUTPUT: Adjust to the most comfortable intensity level.



Freatment Session







REFLEX SYMPATHETIC DYSTROPHY

➢ Setting

MODE: M Mode

PULSE WIDTH: 100 - 150

PULSE RATE: 80 - 100Hz


OUTPUT: Adjust to the most comfortable intensity level.


Treatment Session

24 hours is available for the first 4 days.

4-6 hours daily thereafter.







TRIGEMINAL NEURALGIA

➢ Setting

MODE: M Mode


PULSE WIDTH: 70

PULSE RATE: 100Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session





CERVICAL PAIN

➢ Setting

MODE: C Mode


PULSE WIDTH: 100 - 150


PULSE RATE: 60 - 100Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session







CHRONIC CERVICAL STRAIN

➢ Setting

MODE: M Mode

PULSE WIDTH: 160

PULSE RATE: 30Hz


OUTPUT: Adjust to the most comfortable intensity level.


Treatment Session

24 hours is available for the first 4 days.

4-5 hours daily thereafter.







CHRONIC CERVICAL SPINE PAIN (Postlaminectomy)

➢ Setting

MODE: M Mode

PULSE WIDTH: 200

PULSE RATE: 10Hz


OUTPUT: Adjust to the most comfortable intensity level.


Treatment Session

24 hours is available for the first 4 days.

4-5 hours daily thereafter.







CERVICAL OSTEOARTHRITIS

➢ Setting

MODE: C Mode

PULSE WIDTH: 100 - 150


PULSE RATE: 100Hz


OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available until initial relief.







UNILATERAL CERVICAL SPINE PAIN

➢ Setting

MODE: M Mode

PULSE WIDTH: 100


PULSE RATE: 100Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available until initial relief.





MASECTOMY – RIGHT SIDE

➢ Setting

MODE: M Mode

PULSE WIDTH: 260


PULSE RATE: 120Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available until initial relief.





DEGENERATIVE ARTHRITIS: CERVICAL AND LUMBAR

➢ Setting

MODE: C Mode

PULSE WIDTH: 100


PULSE RATE: 100Hz


OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available until initial relief.







LATERAL RIB CAGE PAIN

➢ Setting

MODE: C Mode


PULSE WIDTH: 150

PULSE RATE: 100Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session





CHRONIC HIP PAIN

➢ Setting

MODE: M Mode


PULSE WIDTH: 200


PULSE RATE: 100Hz


OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session









HERPES ZOSTER NEURALGIA

➢ Setting

MODE: C Mode

PULSE WIDTH: 150


PULSE RATE: 100Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available until initial relief.





Place one set of electrodes on medial side and one set on lateral side

ACUTE MUSCLE AND LIGAMENT TEAR - ANKLE

➢ Setting

MODE: C Mode

PULSE WIDTH: 100


PULSE RATE: 100Hz


OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available until initial relief.







POST-PODIATRIC SURGERY (involving lateral toes)

> Setting

MODE: C Mode

PULSE WIDTH: 100 - 150

PULSE RATE: 100Hz


OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available for the first 4 days.

4 hours daily thereafter.





KNEE PAIN – POST-OP

➢ Setting

MODE: M Mode


PULSE WIDTH: 100 - 150

PULSE RATE: 120Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session





DEGENERATIVE ARTHRITIS - KNEE PAIN

➢ Setting

MODE: C Mode

PULSE WIDTH: 220


PULSE RATE: 80Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available until initial relief.





Primary Placement

Place one set of electrodes on medial side and one set on lateral side

RECURRENT PATELLAR SUBLAXATION

➢ Setting

MODE: C Mode

PULSE WIDTH: 220

PULSE RATE: 80Hz


OUTPUT: Adjust to the most comfortable intensity level.


Treatment Session

24 hours is available for the first 4 days.

4 hours daily thereafter.







Alternate Placement Place one set on each side of leg.

LOW EXTREMITY PAIN (REFLEX SYMPATHETIC DYSTROPHY)

➢ Setting

MODE: C Mode or M Mode

PULSE WIDTH: 160


PULSE RATE: 30 - 80Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available until initial relief.





LOWER LEG PAIN (DIABETIC NEUROPATHY)

➢ Setting

MODE: M Mode


PULSE WIDTH: 100 - 160


PULSE RATE: 60 - 100Hz


OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session









CARPAL TUNNEL SYNDROME

➢ Setting

MODE: C Mode

PULSE WIDTH: 260

PULSE RATE: 100Hz


OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available until initial relief.


20 minutes, 3 times daily thereafter.








Alternate Placement


WRIST PAIN











➢ Setting

MODE: C Mode


PULSE WIDTH: 260


PULSE RATE: 30 - 50Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session







Primary Placement

ELBOW & FOREARM PAIN

➢ Setting

MODE: C Mode


PULSE WIDTH: 100


PULSE RATE: 100Hz

OUTPUT: Adjust to the most comfortable intensity level.


Treatment Session








Primary Placement





UPPER EXTREMITY PAIN (REFLEX SYMPATHETIC DYSTROPHY)

➢ Setting

MODE: C Mode or M Mode

PULSE WIDTH: 220

PULSE RATE: 30 - 50Hz


OUTPUT: Adjust to the most comfortable intensity level.

SUGGESTION: Initial treatment begins with a low pulse width. Look for reduction in swelling and temperature.

Treatment Session

24 hours is available until initial relief.





Primary Placement

ULNAR NERVE LESION

➢ Setting

MODE: C Mode


PULSE WIDTH: 100

PULSE RATE: 100Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session







➢ Setting

MODE: M Mode

PULSE WIDTH: 260


PULSE RATE: 100Hz

OUTPUT: Adjust to the most comfortable intensity level.

Treatment Session

24 hours is available until initial relief.30 minutes, 3 times daily thereafter.





Alternate Placement

Electrode Placement and Charts

The importance of accurate electrode placement

Proper electrode placement is vital to obtain the best results from your TENS unit. Please see the following charts which have been specially prepared to help you.

The most important aspect of TENS electrode placement is to position them so that the current passes through the painful area, or along the nerves leading from the pain.

How to use the electrode placement charts

Check the index on the opposite page for the chart most applicable to the pain problem. We have selected the easiest and most frequently used electrode placement positions and settings. Note: These are suggested settings. There are alternative settings and all settings should be instructed by a medical practitioner.



Low Back Pain Hip Neuralgia Phantom Limb, lower extremity Sciatica **Bicipital Tendonitis** Temporal Mandibular Joint Pain (TMJ) Shoulder Pain **Reflex Sympathetic Dystrophy Trigeminal Neuralgia Cervical Pain** Chronic Cervical Strain **Chronic Cervical Spine Pain (Postlaminectomy) Cervical Osteoarthritis Unilateral Cervical Spine Pain** Masectomy – Right side Degenerative Arthritis: Cervical and Lumbar Lateral Rib Cage Pain **Chronic Hip Pain** Herpes Zoster Neuralgia Acute Muscle and Ligament Tear – Ankle Post Podiatric Surgery (involving lateral toes) Knee Pain (Post-op) **Degenerative Arthritis – Knee Pain Recurrent Patellar Sublaxation** Low Extremity Pain (Reflex Sympathetic Dystrophy) Lower Leg Pain (Diabetic Neuropathy) **Carpal Tunnel Syndrome** Wrist Pain Elbow & Forearm Pain Upper Extremity Pain (Reflex Sympathetic Dystrophy) Ulnar Nerve Lesion **Atypical Facial Pain**